



Krautkramer Transducer Certificate of Conformity Compliant to ASTM E-1065 Guidelines

Date:

2/15/2008

Product Code:

113-135-701

Serial Number:

01RYMD

Description:

FDCP--XDCR FH2EWR 7.5X.3"

Test Setup--

Test Target:


.4" 303 STEEL

Energy:

2

Impedance:

50 Ω



Waveform is the first Return Echo from the indicated target.

Spectrum

UTA S/N:

M00114,

UTA Cal Due Date:

8/9/2008

O-scope SN:


B010911

O-scope Cal Due Date:

10/31/2008

Software:

FNT0040 Rev: H



Test Data--

Sensitivity(Abs db):

13.00

dB

Pulse Duration@-20dB: 6.67E-07

Sec

Peak Frequency:

7.81E+06

Hz

Center Frequency:

7.45E+06

Ηz

% Bandwidth@-6dB:

%

inspector:

CY

29.0

The Ultrasonic Transducer listed in this document has been performance tested and meets all manufacturing specifications. It performed as designed and specified on the applicable style of

Krautkramer instrumentation.

The accuracy of the transducer described above has been confirmed by factory standard test equipment and laboratory reference standards traceable to the National Institute of Standards and Technology. This facility's Quality System is registered to ISO 9001-2000, and is compliant to MIL-STD-45662A and ANSI/NCSL Z540-1-1994.



GE Inspection Technologies, LP 50 Industrial Park Rd. Lewistown, PA 17044

Tel.: 717.242.0327 Fax: 717.242.2606

GEInspectionTechnologies.com