


Transducer Certificate of Conformity	
Part Number	113-292-604
Description	XDCR,AWS,GAMMA,2.25,.75"SQ
Serial Number	24E021MM
Test Result	Pass
Test Date	06/11/2024
Inspector	JG
Test Procedure	0511044



Test Data				
Test Parameter	Units	Lower Spec Lim.	Measured	Upper Spec Lim.
Loop Sensitivity	dB	NA	-44.34	NA
Center Frequency	MHz	2.025	2.28	2.475
Peak Frequency	MHz	2	2.17	2.5
Bandwidth @ -6dB	%	32	39	NA
20dB Ringdown	µs	NA	1.88	2.3
JSR Gain	dB	NA	-4.5	-1

Calibrated Test Equipment			
Item	Description	Serial No.	Cal. Due
Pulser	JSR DPR300	DF0102	12/06/2024
Oscilloscope	Tektronix TDS2012C	C051687	01/17/2025
Target	1.0" PLEXI GLASS	1289/20	2/1/2025
Func. Generator	NA	NA	NA

JSR Pulser Settings						
Rec. Imp.	Pul. Imp.	Energy	Voltage	Mode	HP Filter	LP Filter
101 Ohms	LOW	MID-LOW	300	T/R	1.0 MHz	22.5 MHz

Declaration:

The Ultrasonic Transducer in this document has been performance tested and meets all manufacturing specifications.

We certify that the product mentioned was subjected to a quality inspection and testing procedure. It conforms to our manufacturing drawings, meets our specifications and compliant to ASTM E-1065 guidelines. All measuring tools used for testing are calibrated and traceable to national or international standards. Our quality management system is certified according to EN ISO 9001.

Baker Hughes Slovakia. s.r.o
 Trencianska 1877/15
 Nove Mesto nad Vahom
 Slovakia, 91528

For Customer Service, see "Contact us" on webpage:
www.bakerhughesds.com/waygate-technologies

