



www.llogsa.com

GE Inspection Technologies

Transducer Certificate of Conformity Compliant to ASTM E-1065 Guidelines

Date: 9/25/2014
Product Code: 113-844-600
Serial Number: 14I01GZZ
Description: FABP--XDCR SWS 5X.5 BNC

Test Setup--
Test Target: .75" PLEXI GLASS
Energy: 2
Impedance: 100 Ω


UTA S/N: MM00108
UTA Cal Due Date: 4/18/2015
O-scope SN: C020916
O-scope Cal Due Date: 10/31/2014

Software: FNT0040 Rev: H


Test Data--
Sensitivity(Abs db): -9.40 dB

Pulse Duration@-20dB: 4.67E-07 Sec
Peak Frequency: 5.09E+06 Hz
Center Frequency: 4.25E+06 Hz
% Bandwidth@-6dB: 122.0 %

Inspector: JPT



Waveform is the first Return Echo from the indicated target.



The Ultrasonic Transducer listed in this document has been performance tested and meets all manufacturing specifications. It performed as designed and specified on the applicable style of Krautkramer instrumentation.

The accuracy of the transducer described above has been confirmed by factory standard test equipment and laboratory reference standards traceable to the National Institute of Standards and Technology. This facility's Quality System is registered to ISO 9001-2008, and is compliant to MIL-STD-45662A and ANSI/NCSL Z540-1-1994.

GE Inspection Technologies, LP
50 Industrial Park Rd.
Lewistown, PA 17044
Tel.: 717.242.0327
Fax: 717.242.2606
GEInspectionTechnologies.com