



Krautkramer Transducer Certification

Compliant to ASTM E-1065 Guidelines

Date: 3/20/2012
Product Code: 389-066-020
Serial Number: 025D57
Description: PAB2B2.25/45SCS/.575ODCSE/S


Test Setup--
Test Target: 1" B1 Block
Energy: 2
Impedance: 50 Ohms

Equipment Used:
UTA Pulser: UTA5
UTA S/N: MM00114
UTA Cal Due Date: 9/21/2012
O-scope SN: B012087
O-scope Cal Due Date: 10/31/2012


Software: FNT0040 **Rev:** H

Test Data--
Sensitivity(Abs db): 15.60 dB
Pulse Duration@-20dB: 9.80E-07 Sec
Peak Frequency: 2.39E+06 Hz
Center Frequency: 2.56E+06 Hz
% Bandwidth@-6dB: 66.0 %

Inspector: CY



Waveform is the first Return Echo from the indicated target.



The accuracy of the transducer described above has been confirmed by factory standard test equipment and laboratory reference standards traceable to the National Institute of Standards and Technology. The uncertainty factor of the block standards used does not exceed 25% of the tolerance of the characteristic being verified. This facility's Quality System is registered to ISO 9001-2008, and is compliant to MIL-STD-45662A and ANSI/NCSL Z540-1-1994.

GE Inspection Technologies, LP
 50 Industrial Park Rd.
 Lewistown, PA 17044
 Tel.: 717.242.0327
 Fax: 717.242.2606
 GEInspectionTechnologies.com